

REMARKS

Applicants respectfully requests reconsideration of the present application in view of the following reasons.

I. Status of the Claims

No amendments are made in this response. Claims 28-36, 39, 40, 42, 43, 51-60, and 64-74 are pending and under examination.

II. Rejection of Claims under 35 U.S.C. §103(a)

A. Liversidge, Pavord, Glaxo History, Merck Index, Radhakrishnan and Palmer

Claims 28-36, 39-40, 51-60 and 64-73 remain rejected under 35 U.S.C. §103(a) for allegedly being obvious over U.S. Patent No. 5,145,684 to Liversidge et al. (“Liversidge”) in view of Pavord et al., *Clin. Pharmacokinet.*, 25(2) (1993), abstract (“Pavord”), Glaxo History (www.gsk.com/about/history-noflash.htm), and the Merck Index, 10th ed., page 144, entry 1018 (“Merck Index”), as evidenced by U.S. Patent No. 5,049,389 to Radhakrishnan et al. (“Radhakrishnan”) and U.S. Patent No. 5,208,226 to Palmer et al. (“Palmer”). Applicants respectfully traverse the rejection.

According to the Examiner, the claimed invention is unpatentable over the cited combination of references because it would have been obvious to reduce the particle size of beclomethasone, a drug listed in the Merck Index for treating asthma, in view of the teaching of Liversidge, and making the nanoparticulate beclomethasone composition into an aerosol formulation in view of the teaching of Glaxo History. Applicants respectfully submit that a *prima facie* case of obviousness is not established because (1) there is unpredictability in the art of making nanoparticulate active agent compositions such that one of ordinary skill in the art would not have been motivated to combine the cited references with any reasonable expectation

of success, and (2) the Bosch Declaration properly provides evidence to compare the claimed invention with the closest art.

(1) It was unpredictable that a nanoparticulate beclomethasone composition could be obtained.

Applicants submit herewith a declaration under 37 C.F.R. 1.132 executed by Dr. Gary G. Liversidge (“the Liversidge Declaration”) to demonstrate that not all active agents can be made into stable nanoparticulate active agent compositions. Particularly, it is unpredictable whether a functional equivalent can be successfully made into a nanoparticulate active agent composition. *See* ¶¶ 4-24.

The Examiner contends that “Liversidge explicitly identifies corticosteroids and steroids” as exemplary active agents for the nanoparticulate active agent compositions, and therefore it would have been obvious to use Liversidge’s teaching to obtain a nanoparticulate beclomethasone active agent composition. *See* Office Action, page 12, 1st full paragraph.

As Dr. Liversidge testifies, it is unpredictable that an active agent can be successfully made into a nanoparticulate active agent composition even if a nanoparticulate composition comprising another active agent having the same function was successfully obtained. *See* the Liversidge Declaration, ¶¶ 4-24.

Accordingly, the skilled artisan would *not* have considered it obvious to obtain a nanoparticulate beclomethasone composition in view of the combined teachings of the cited references.

(2) Liversidge is incorrectly identified as the “closest art.”

The Examiner dismissed the evidence submitted in the Bosch Declaration asserting that the comparison between the aerosol formulation comprising a nanoparticulate beclomethasone composition of the claimed invention was “an apples and oranges comparison regarding the

deposition of the compositions contained in the two different devices. Thus, the Examiner requires that Applicants compare the claimed invention with Liversidge. *See* Office Action, the paragraph bridging pages 3 and 4, and at page 4, 1st full paragraph.

MPEP 716.02(e) explicitly sets forth that the “closest art” does not necessarily have to be the reference identified by the Examiner. Rather, “Applicants may compare the claimed invention with prior art that is more closely related to the invention than the prior art relied upon by the examiner. *In re Holladay*, 584 F.2d 384, 199 USPQ 516 (CCPA 1978).”

In the present case, Applicants respectfully disagree that Liversidge is the closest art. This is because Liversidge, as acknowledged by the Examiner, fails to explicitly disclose a nanoparticulate active agent composition ***having beclomethasone as the active agent***. Additionally, Liversidge ***fails to disclose an aerosol formulation*** comprising the nanoparticulate active agent composition. For at least these reasons alone, the comparison between the aerosol formulation comprising nanoparticulate beclomethasone composition and Liversidge would be an apple-to-orange comparison in terms of comparing different active agents **and** different dosage forms.

In contrast, VANCERIL[®] is the commercial aerosol formulation comprising microparticulate beclomethasone as the active agent – e.g., the ***same*** active agent and the ***same*** dosage form as the claimed invention. Comparison of VANCERIL[®] with Applicants’ aerosol formulation comprising a nanoparticulate beclomethasone composition requires a comparison of the same active agent, as well as the same dosage form. As such, VANCERIL[®] rather than Liversidge is the closest art for comparison purposes to demonstrate unexpected results.

In this regard, the Examiner asserts an “apple-and-orange” comparison between VANCERIL[®] and the claimed invention merely based on the difference in the dosing device. This assertion does not have any valid basis in view of the commonalities of active agent and dosage form, which are lacking for the comparison between Liversidge and the claimed

invention, required by the Examiner. Moreover, the Bosch Declaration attests to the fact that an ultrasonic nebulizer is efficient in delivering solutions but not water-insoluble active agents. *See* ¶10. *See also* Hess, "Nebulizers: Principles and Performance," *Respiratory Care*, 45(6): 609-622 (2000) (Exhibit 1). Exhibit 1 discusses nebulizing solutions using an ultrasonic nebulizer (page 618, left column, 2nd full paragraph; and right column, 1st full paragraph). Furthermore, Exhibit 1 discloses that disadvantages of an ultrasonic nebulizer include the fact that not all drug formulations are available for use in an ultrasonic nebulizer and the drug preparation required for use in an ultrasonic nebulizer. *See* page 618, Table 4. Additionally, the factors affecting the output from ultrasonic nebulizers include fluid density, viscosity, surface tension and vapor pressure, etc. *Id.*, Table 5.

Because of the claimed invention, a nanoparticulate beclomethasone composition, due to its reduced particle size, can now be delivered by an ultrasonic nebulizer with improved efficiency as compared to prior art VANCERIL[®] formulation. *See* ¶¶12-17.

Finally, when a jet nebulizer was used (e.g., the same type of nebulizer used for delivering both nanoparticulate and microparticulate formulations), nanoparticulate beclomethasone formulations were delivered with higher efficiency than a microparticulate beclomethasone formulation. *See* the Bosch Declaration, ¶¶4-8. In this comparison, the same active agent, same type of dosage form, and same delivery device but different particle size were compared.

(3) The Bosch Declaration is commensurate in scope with the claimed invention.

The Examiner asserts that the showing in the Bosch Declaration is not commensurate in scope with the claimed invention because except for claims 73-74, Applicants' claims are not limited to any particular surface stabilizer.

Pursuant to MPEP 716.02(d), “[t]he nonobviousness of a broader claimed range can be supported by evidence based on unexpected results from testing a narrower range if one of ordinary skill in the art would be able to determine a trend in the exemplified data which would allow the artisan to reasonably extend the probative value thereof.” In the present case, the Bosch Declaration demonstrated improved delivery efficiency associated with a nanoparticulate beclomethasone composition, *i.e.*, the reduced particle size of the active agent. Therefore, the skilled artisan would have been able to determine the trend in the correlation with reduced particle size and employ routine experimentation to identify suitable surface stabilizers to maintain the small particle size of beclomethasone. Accordingly, the evidence presented in the Bosch Declaration is commensurate in scope with the claimed invention.

For the same reason, claims 73 and 74 benefit from additional grounds of patentability because unexpected results were achieved by aerosol formulations comprising nanoparticulate beclomethasone compositions having polyvinyl alcohol or tyloxapol as a surface stabilizer.

B. Liversidge, Pavord, Glaxo History, Merck Index, Radhakrishnan, Palmer and Spear

Claims 42-43 are rejected under 35 U.S.C. §103(a) for allegedly being obvious over Liversidge in view of Pavord, Glaxo History, and the Merck Index, as evidenced by Radhakrishnan and Palmer, and further in view of U.S. Patent No. 5,525,623 to Spear et al. (“Spear”). Applicants respectfully traverse the rejection.

Spear is cited for the alleged teaching of a jet nebulizer or an ultrasonic nebulizer. The Examiner asserts that Applicants attacked the references individually by arguing that Spear shows delivery of a *solution* using a jet nebulizer or an ultrasonic nebulizer but fails to teach using a jet nebulizer or an ultrasonic nebulizer to deliver *a poorly-soluble active agent* or a composition in which the active agent existed in *a particulate form*. Applicants respectfully disagree.

The burden is on the Examiner to show a reason for the skilled artisan to combine the teachings of the cited references. Applicants merely demonstrate that there is no reason that the skilled artisan would have applied the teaching of Spear, which is directed to nebulizing solutions, to nebulizing active agent in particulate form. As discussed above, Exhibit 1 demonstrates that an ultrasonic nebulizer is used for delivery of solutions, and that an ultrasonic nebulizer has the disadvantages of not being available for delivering all drug formulations and requiring drug preparation, and the factors affecting output from an ultrasonic nebulizer include fluid density, viscosity, etc. Spear fails to teach that an ultrasonic nebulizer is suitable for delivering a drug in a particulate form, such as that required by Applicants' claims. Therefore, the Examiner has failed to establish a *prima facie* case of obviousness by articulating a reason to combine the cited references.

C. Liversidge, Pavord, Glaxo History, Merck Index, Radhakrishnan, Palmer and June

Claim 74 is rejected under 35 U.S.C. §103(a) for allegedly being obvious over Liversidge in view of Pavord, Glaxo History, and the Merck Index, as evidenced by Radhakrishnan and Palmer, and further in view of EP 0602701 to June et al. ("June"). Applicants respectfully traverse the rejection.

June is cited for the alleged teaching of tyloxapol as a surface stabilizer, dispersant and wetting agent. *See* Office Action, page 16, 3rd paragraph. However, June fails to compensate for the deficiencies of the remaining references, as discussed in Section II A (1) above. Additionally, the Bosch Declaration demonstrated the unexpected results achieved by a nanoparticulate beclomethasone composition comprising tyloxapol as the surface stabilizer.

In view of the foregoing, Applicants respectfully request withdrawal of the rejection under 35 U.S.C. §103(a).

III. Provisional Double Patenting Rejection

Claims 28-33, 39-40, 51-60, 66, 69, 72, and 74 are provisionally rejected on the ground of nonstatutory obviousness-type double patenting over claims 1-7, 9-11, and 13-14 of copending Application No. 10/035,324 entitled “Sterile Filtered Nanoparticulate Formulations Of Budesonide And Beclomethasone Having Tyloxapol As A Surface Stabilizer” (“the ‘324 application”) in view of Liversidge and Radhakrishnan.

Additionally, claims 28-36 and 51-60 remain provisionally rejected on the ground of nonstatutory obviousness-type double patenting over claims 1-11 and 17-18 of copending Application No. 12/292,092 entitled “Nanoparticulate Compositions Of Immunosuppressive Agents” (“the ‘092 application”) in view of Liversidge and Radhakrishnan. Applicants respectfully traverse each rejection.

Both the ‘324 application and the ‘092 application were filed after the present application. Pursuant to MPEP 804, “[i]f a ‘provisional’ nonstatutory obviousness-type double patenting (ODP) rejection is the only rejection remaining in the earlier filed of the two pending applications, while the later-filed application is rejectable on other grounds, the examiner should withdraw that rejection and permit the earlier-filed application to issue as a patent without a terminal disclaimer.” Accordingly, should the Examiner find the pending claims are allowable, Applicants respectfully request withdrawal of the provisional double patenting rejection without filing a terminal disclaimer in the present application.

CONCLUSION

The present application is now in condition for allowance. Favorable reconsideration of the application as amended is respectfully requested. The Examiner is invited to contact the undersigned by telephone if it is felt that a telephone interview would advance the prosecution of the present application.

The Commissioner is hereby authorized to charge any additional fees which may be required regarding this application under 37 C.F.R. §§ 1.16-1.17, or credit any overpayment, to Deposit Account No. 19-0741. Should no proper payment be enclosed herewith, as by the credit card payment instructions in EFS-Web being incorrect or absent, resulting in a rejected or incorrect credit card transaction, the Commissioner is authorized to charge the unpaid amount to Deposit Account No. 19-0741. If any extensions of time are needed for timely acceptance of papers submitted herewith, Applicants hereby petition for such extension under 37 C.F.R. §1.136 and authorizes payment of any such extensions fees to Deposit Account No. 19-0741.

Respectfully submitted,

Date: February 10, 2012

By 

FOLEY & LARDNER LLP
Customer Number: 31049
Telephone: (202) 672-5538
Facsimile: (202) 672-5399

Michele M. Simkin
Attorney for Applicant
Registration No. 34,717

EXHIBIT 1

Nebulizers: Principles and Performance

Dean R Hess PhD RRT FAARC

Introduction

Pneumatic Nebulizers

Principle of Operation

Clinically Important Characteristics of Nebulizer Performance

Technical Factors Affecting Nebulizer Performance

Patient Factors Affecting Nebulizer Performance

Designs to Enhance Nebulizer Performance

Use of Reservoir Bags to Collect Aerosol During the Expiratory Phase

Breath-Enhanced Nebulizers

Breath-Actuated Nebulizers

Continuous Nebulization

Nebulizers for Specific Applications

Ultrasonic Nebulizers

Summary

[Respir Care 2000;45(6):609–622] *Key words: nebulizer, pneumatic nebulizer, ultrasonic nebulizer, continuous nebulization, aerosol therapy.*

Introduction

Nebulizers are used to convert liquids into aerosols of a size that can be inhaled into the lower respiratory tract. The process of pneumatically converting a bulk liquid into small droplets is called atomization. Pneumatic nebulizers have baffles incorporated into their design so that most of the droplets delivered to the patient are within the respirable size range of 1–5 μm . Ultrasonic nebulizers use electricity to convert a liquid into respirable droplets.

Although the first choice of aerosol generator for the delivery of bronchodilators and steroids is the metered-dose inhaler,^{1,2} nebulizers remain useful for several reasons. First, some drugs for inhalation are available only in solution form. Second, some patients cannot master the

correct use of metered-dose inhalers or dry powder inhalers. Third, some patients prefer the nebulizer over other aerosol generating devices. The physiologic benefits of metered-dose inhalers and nebulizers are virtually equivalent,^{3,4} and the choice of device is often based on clinician or patient preference rather than clear superiority of one approach over the other. Although cost savings have been suggested with the use of metered-dose inhalers compared to nebulizers, these benefits may be overestimated.³

The purpose of this paper is to review the performance characteristics of nebulizers. Both pneumatic and ultrasonic nebulizer designs will be considered.

Pneumatic Nebulizers

Nebulizers are the oldest form of aerosol generation. Although they have been commonly used for many years, their basic design and performance has changed little over the past 25 years. Nebulizers are most commonly used for bronchodilator administration, and it is well established that nebulized bronchodilators produce a physiologic response. Because bronchodilators are relatively inexpensive, there is little market pressure to improve nebulizer performance. In fact, the market generally prefers an inexpensive nebulizer rather than a high-performance neb-

Dean R Hess PhD RRT FAARC is affiliated with Massachusetts General Hospital and Harvard Medical School, Boston, Massachusetts.

A version of this paper was presented by Dr Hess during the RESPIRATORY CARE Journal Consensus Conference, "Aerosols and Delivery Devices," held September 24–26, 1999 in Bermuda.

Correspondence: Dean R Hess PhD RRT FAARC, Respiratory Care, Ellison 401, Massachusetts General Hospital, 55 Fruit Street, Boston MA 02114. E-mail: dhess@partners.org

Table 1. Factors Affecting Penetration and Deposition of Therapeutic Aerosols Delivered via Jet Nebulizer

Technical Factors

Manufacturer of nebulizer
Gas flow used to power nebulizer
Fill volume of nebulizer
Solution characteristics
Composition of the driving gas
Designs to enhance nebulizer output
Continuous versus breath-actuated

Patient Factors

Breathing pattern
Nose versus mouth breathing
Composition of inspired gas
Airway obstruction
Positive pressure delivery
Artificial airway and mechanical ventilation

ulizer for bronchodilator administration. However, there are newer drugs available for inhalation that are expensive and for which precise dosing may be important. These include dornase alfa, tobramycin, and pentamidine. Nebulizer performance is affected by both technical and patient-related factors (Table 1).

Principle of Operation

The operation of a pneumatic nebulizer requires a pressurized gas supply as the driving force for liquid atomization (Fig. 1).⁵⁻¹⁰ Compressed gas is delivered through a jet, causing a region of negative pressure. The solution to be aerosolized is entrained into the gas stream and is sheared into a liquid film. This film is unstable and breaks into droplets because of surface tension forces. A baffle is placed in the aerosol stream, producing smaller particles and causing larger particles to return to the liquid reservoir. More than 99% of the particles may be returned to the liquid reservoir.⁸ The aerosol is delivered into the inspiratory gas stream of the patient. Before delivery into the patient's respiratory tract, the aerosol can be further con-




Fig. 1. Basic components of the design of pneumatic nebulizers. (Adapted from Reference 6.)




Fig. 2. Internal mixing and external mixing nebulizer designs. (Adapted from Reference 5.)

ditioned by environmental factors such as the relative humidity of the carrier gas.¹¹⁻¹⁴

Nebulizer nozzles are of two types (Fig. 2).¹⁵ With the internal mixing design, gas flow interacts with the solution prior to leaving the exit port. With external mixing, gas and the solution interact after both leave the nozzle. Modifications on these designs are used by nebulizer manufacturers, without clear superiority of one approach over the other.

Determinants of droplet size produced by nebulizers include the characteristics of the solution (density, viscosity, surface tension), the velocities of the gas and solution, and the flow rates for the gas and the solution.^{5,15} The most important factors are gas velocity and the ratio of liquid to gas flow.⁵ An increase in gas velocity decreases droplet size, whereas an increase in the ratio of liquid to gas flow increases particle size. It is interesting to note that gas velocity affects the flow rates for both the gas and the solution. Thus, it is impossible to separately control the primary factors affecting droplet size from nebulizers.

An important consideration in the use of nebulizers is the dead volume of the device. Dead volume refers to the amount of solution that is trapped inside the nebulizer and is thus not made available for inhalation. The dead volume is typically in the range of 1 to 3 mL. Dead volume is minimized by using a conical shape of the nebulizer, by decreasing the surface area of the internal surface of the nebulizer, and by improving the wetness of the plastic surface of the nebulizer.^{5,15} To reduce medication loss due to dead volume, clinicians and patients may tap the nebulizer periodically during therapy, which has been shown to increase nebulizer output.¹⁶ Therapy may also be continued past the point of inconsistent nebulization (sputtering) in an attempt to deliver medication from the dead volume, but this has been reported to be unproductive.¹⁷




Fig. 3. Comparison of the output of 17 commercially available pneumatic nebulizers. Respirable mass = particles 1–5 μm delivered to mouthpiece with simulated spontaneous breathing; 2.5 mg albuterol placed into nebulizer cup. (Adapted from Reference 25.)

Due to evaporative water loss, the solution in the nebulizer becomes increasingly concentrated during the nebulization time.^{18–20} For this reason, gravimetric methods underestimate nebulizer output.^{21,22} Due to the evaporative effects, the nebulizer solution cools.¹⁸ Solution temperature affects nebulizer output, with output varying directly with temperature.^{15,23} The droplet size produced by the nebulizer also varies directly with temperature.¹⁸

Clinically Important Characteristics of Nebulizer Performance

The most important characteristic of nebulizer performance is the respirable dose provided for the patient. The respirable dose is determined by the mass output of the nebulizer and the size of the droplets that are produced. The droplet size should be 2–5 μm for airway deposition and 1–2 μm for parenchymal deposition.²⁴ Droplet size is usually reported as mass median aerodynamic diameter (MMAD), which is the diameter around which the mass of the aerosol is equally divided. Note that MMAD is used to characterize the population of droplets produced; it does not refer to the size of individual droplets. Because the volume, and hence the mass, of the droplet is determined by the cube of the radius (volume = $4/3 \pi r^3$), most of the particles will be smaller than the MMAD. The respirable dose is sometimes reported as respirable mass, which is the output of droplets from a nebulizer in the respirable range of 1–5 μm .^{25,26}

Other important characteristics of nebulizer performance include nebulization time, cost, ease of use, and requirements for cleaning and sterilization. Nebulization time is important for patient compliance in the outpatient setting

and clinician supervision for hospitalized patients. A short nebulization time that delivers an effective dose is desirable. Many nebulizers are low-cost, mass-produced, single-patient-use devices. This results in variability in performance among devices,^{25–28} which might not be important for bronchodilator delivery, but which could be important for delivery of other inhaled medications.

Technical Factors Affecting Nebulizer Performance

Several studies have reported performance differences between nebulizers from different manufacturers.^{23,25–27,29–36} Performance differences among nebulizers from the same manufacturer have been reported.^{37–39} Hess et al²⁵ evaluated the performance of 17 nebulizers, using a model of spontaneous breathing. They reported a respirable mass available to the patient that was severalfold greater from some nebulizers than from others (Fig. 3). Performance differences between nebulizers may have clinical implications.^{37,40–43} In healthy subjects, Hardy et al⁴² reported aerosol deposition from some pneumatic nebulizers that was twice that of others. In subjects with chronic stable asthma, Johnson et al⁴³ reported differences in bronchodilation between nebulizers from different manufacturers.

Because of cost considerations, disposable single-patient-use nebulizers are typically used for many treatments. The effects of repetitive use and cleaning were evaluated by Standaert et al.⁴⁴ In that study, nebulizers were found to function correctly for 100 repeated uses, provided they were properly maintained. Proper maintenance consisted of washing with soapy water, rinsing, and air drying after each use. Each nebulizer was also subjected to a daily 30-minute soak in 2.5% acetic acid. In the same study,⁴⁴ it




Fig. 4. Comparison of albuterol output of a pneumatic nebulizer, using two formulations of albuterol. (Drawn from data in Reference 38.)

was reported that the nebulizers started to fail after 40 uses if they were not cleaned after each use.

Several studies have reported greater output from pneumatic nebulizers when the fill volume is increased.^{23,25,33,34,45} This is probably because nebulizers have a fixed dead volume, and thus an increase in fill volume reduces the proportion of dead volume within the nebulizer. Although nebulizer output increases with a greater fill volume, there is also an increase in nebulization time.²⁵ The nebulization time can be reduced when a larger fill volume is used by increasing the flow to power the nebulizer.^{23,25} A nebulizer fill volume of 4–5 mL is recommended.

Output increases with an increased flow to power the nebulizer.^{23,25,33} An increase in flow also decreases the droplet size produced by nebulizers.^{25,39,46–48} A flow of 8 L/min is recommended. Flows lower than this result in decreased nebulizer performance. A flow greater than this may result in increased drug loss during the expiratory phase, which offsets the effect of greater flow on nebulizer output.²⁵

It is not commonly appreciated that the drug formulation can affect nebulizer performance. MacNeish et al³⁸ reported differences in nebulizer output with two formulations of albuterol. Nebulizer output was significantly greater with the formulation containing the preservative benzalkonium chloride, probably because of its surface activity (Fig. 4). Large droplets were seen to adhere to the walls of the nebulizer with the preservative-free formulation, whereas foaming was seen to occur with the preservative-containing formulation. Others have also reported effects of drug formulation on nebulizer output.^{39,49} It is interesting to note that metered-dose inhalers have always been tested and approved as a drug-delivery-system combination. Newer drug solutions have also been approved

for a specific nebulizer (eg, pentamidine, ribavirin, dornase alpha, tobramycin).

The density of the gas powering the nebulizer affects nebulizer performance. Hess et al⁵⁰ reported the effect of heliox (80% helium, 20% oxygen) on nebulizer function. The inhaled mass of albuterol was significantly reduced when the nebulizer was powered with heliox, and there was a greater than twofold increase in nebulization time with heliox. An increased flow with heliox produced a respirable mass output similar to that produced when the nebulizer was powered with air. These results are explained by Bernoulli's principle, which predicts that the decreased density of heliox increases the velocity at which the gas leaves the jet orifice and produces less negative pressure to entrain drug solution.⁵⁰

Patient Factors Affecting Nebulizer Performance

The breathing pattern of the patient affects the amount of aerosol deposited in the lower respiratory tract. This partially explains differences in aerosol deposition between children and adults. To improve aerosol penetration and deposition in the lungs, the patient should be encouraged to use a slow and deep breathing pattern.⁵¹ Because of the effect of breathing pattern on drug delivery from nebulizers, *in vitro* evaluations of nebulizer performance should be conducted in a manner that simulates the breathing pattern of a patient.⁵²

Inhaled aerosols can be administered using a mouthpiece or a face mask.^{53–57} Bronchodilator responses occur with both techniques, and some have argued that the selection of interface should be based on patient preference.⁵³ However, it should be appreciated that the nasal passages effectively filter droplets delivered from the nebulizer. Everard et al⁵⁴ reported a nearly 50% reduction in aerosol




Fig. 5. Peak expiratory flow (PEF) responses with albuterol delivery by nebulizer with bilevel positive airway pressure (BiPAP) or by nebulizer alone. (Drawn from data in Reference 68.)

delivery to the lungs with nasal inhalation. Whether a mouthpiece or a face mask is used, it is important to instruct the patient to inhale through the mouth. Use of a mouthpiece may encourage oral breathing.⁵⁴ It is interesting to note that asthmatic patients switch their breathing route from the nasal route to the oronasal route during acute exacerbations and, even when not acutely bronchoconstricted, switch to oronasal breathing when wearing a face mask.⁵⁸

Airway caliber affects lung delivery of nebulized bronchodilators.⁵⁹⁻⁶¹ Lipworth et al⁶¹ reported lower plasma albuterol concentrations and attenuated bronchodilator responses in patients with severe asthma than in normal subjects or those with mild asthma. It is ironic that the air flow obstruction that produces the need for inhaled bronchodilator therapy also decreases the effectiveness of that therapy.

Several studies have reported greater pulmonary penetration of aerosols in patients with stable asthma and with acute airway constriction during heliox breathing.⁶²⁻⁶⁶ Because of the lower density and greater viscosity of heliox, gas flow becomes less turbulent, which theoretically improves the transport of aerosols through constricted airways to more peripheral lung regions. Henderson et al⁶⁷ reported no significant advantage of heliox-driven nebulizer therapy over oxygen-driven nebulizer therapy. However, they⁶⁷ did not account for the effect of heliox on nebulizer function,⁵⁰ and this may have contributed to their negative findings.

Nebulizer therapy has been used in combination with noninvasive ventilation. Pollack et al⁶⁸ randomized patients with acute asthma to receive either bronchodilator therapy with a nebulizer and face mask or with a nebulizer, nasal mask, and BiPAP (bi-level positive airway pressure system made by Respirationics). The BiPAP settings were: inspiratory pressure 10 cm H₂O, expiratory pressure 5 cm

H₂O. The patients who received bronchodilator therapy with BiPAP had a greater improvement in peak flow (Fig. 5). Although these results are intriguing, further positive reports are needed before widespread acceptance of this practice. Interestingly, this approach to nebulizer therapy is reminiscent of intermittent positive-pressure breathing, which was abandoned many years ago as a method for delivery of inhaled bronchodilators to patients with asthma.

Nebulizer therapy is commonly used in mechanically ventilated patients, and this topic has been reviewed in detail elsewhere.⁶⁹⁻⁷³ A number of factors are known to affect aerosol delivery from nebulizers during mechanical ventilation (Table 2).¹² There are disadvantages of nebulizer use during mechanical ventilation, such as circuit contamination,⁷⁴ decreased ability of the patient to trigger the ventilator⁷⁵ (if the nebulizer is not powered by the ventilator), and increases in the delivered tidal volume and airway pressure⁷⁶ (if the nebulizer is not powered by the ventilator). The nebulizer is less efficient than the metered-dose inhaler during mechanical ventilation, but the nebulizer delivers a greater dose to the lower respiratory tract.⁷⁷

Designs to Enhance Nebulizer Performance

In recent years, several nebulizer designs have become available to decrease the amount of aerosol lost during the expiratory phase.⁷⁸ These include reservoir bags to collect aerosol during the expiratory phase, the use of a vented design to increase the nebulizer output during the inspiratory phase (breath-enhanced nebulizers), and nebulizers that only generate aerosol during the inspiratory phase (breath-actuated nebulizers). Because these designs improve drug delivery to the patient, they have the potential to reduce treatment time, which should improve patient compliance with nebulizer therapy.

Use of Reservoir Bags to Collect Aerosol During the Expiratory Phase

For many years, it has been a common practice to use a T-piece and corrugated tubing as a reservoir for small-

Table 2. Factors Affecting Aerosol Delivery from Nebulizers During Mechanical Ventilation

Endotracheal tube size
Position of nebulizer placement in the circuit
Type of nebulizer and fill volume
Humidification of the inspired gas
Treatment time
Duty cycle (I:E ratio)
Ventilator brand

I:E ratio = ratio of inspiratory time to expiratory time.




Fig. 6. Two designs of nebulizer device that use a reservoir bag. Circulaire (left) and AeroTee (right).

volume nebulizers.⁷⁹ In the late 1980s and early 1990s, there were reports of increased aerosol delivery to the lower respiratory tract when a plastic chamber was used with the nebulizer to capture aerosol during the expiratory phase, and provide that to the patient during the subsequent inspiration.^{80,81} In the United States, a similar concept was incorporated into the Circulaire and AeroTee designs (Fig. 6). Both of these designs use a 750 mL bag to store aerosol during exhalation, but differ in how they prevent rebreathing. The Circulaire uses a one-way valve to prevent exhaled gas from entering the reservoir bag, whereas the AeroTee allows some exhaled gas to enter the bag.⁸² These designs also decrease environmental contamination with the aerosol that is generated. The Circulaire incorporates a variable inspiratory/expiratory resistor that is set to maximize inspiration from the reservoir bag, and to provide a positive expiratory pressure effect.

Mason et al⁸³ reported an MMAD of 0.51 μm with the Circulaire. Compared with a conventional nebulizer, they also reported better lung deposition, less gastrointestinal deposition, and less drug loss to the environment (Fig. 7). However, there are several important observations about these results. First, of 9 normal subjects, 2 actually had decreased pulmonary deposition with the Circulaire. This illustrates why caution must be exercised when applying group data to individual patients. Second, the conventional nebulizer used by Mason et al⁸³ does not perform as well as the nebulizer incorporated into the Circulaire. Thus, it is unclear whether the results were the effect of the reservoir bag or the nebulizer. The MMAD reported by Mason et al⁸³ for the Circulaire is not ideal. For maximal pulmonary deposition of bronchodilators, an MMAD of 1–5 μm is more desirable.

In another study by Mason et al,⁸⁴ the Circulaire was compared to a conventional nebulizer for bronchodilator delivery in patients with chronic obstructive pulmonary disease. In that study, the pulmonary deposition and therapeutic effect were similar for the Circulaire and the conventional nebulizer. Hoffman et al⁸⁵ compared the Circulaire to a conventional nebulizer for bronchodilator delivery in patients with acute bronchospasm presenting to an emergency department. They reported a greater improvement in

bronchospasm (measured by peak flow) in the Circulaire group. In this study,⁸⁵ like those by Mason et al,^{83,84} the nebulizer used with the Circulaire may have been superior to the conventional nebulizer that was used and, thus, the study may have compared the performance of nebulizers rather than the effect of the reservoir bag.

Breath-Enhanced Nebulizers

The traditional nebulizer design incorporates the nebulizer sidestream to the air flow of the patient. Some newer nebulizers use a mainstream design with valves. In this valved open-vent design, the patient breathes through the nebulizer during inspiration, which enhances the nebulizer output. During the expiratory phase, a one-way valve directs patient flow away from the nebulizer chamber (Fig. 8).

This design has been evaluated in several studies, which have reported greater pulmonary deposition with this design than with a conventional nebulizer.^{86–88} A potential advantage of the open-vent nebulizer design is an improvement in nebulizer output with an increase in inspiratory flow. Coates et al³¹ reported a greater output of tobramycin with increased inspiratory flow, using an open-vent nebulizer, whereas changes in inspiratory flow did not affect the output of the conventional nebulizer (Fig. 9). As with conventional nebulizers, performance differences between breath-enhanced nebulizers have been reported.^{88,89}

Breath-Actuated Nebulizers

Aerosol waste during the expiratory phase can be eliminated if the nebulizer is only active during the inspiratory phase. Methods to manually actuate the nebulizer during the inspiratory phase have been available for many years.^{45,90} It is also of interest to note that this design is commonly used in mechanical ventilator-actuated designs.^{91,92} Both pneumatically and electronically controlled breath-actuated nebulizers have recently become commercially available. Their role in clinical application is yet to be determined.

Continuous Nebulization

Since the late 1980s, there has been considerable clinical and academic interest in the use of continuous aerosolized bronchodilators for the treatment of acute asthma^{93–106} (Table 3). These studies suggest that this therapy is safe, at least as effective as intermittent nebulization, and may be superior to intermittent nebulization in patients with the most severe pulmonary function.

Several configurations have been described for continuous nebulization.¹⁰⁷ These include frequent refilling of




Fig. 7. Percent of drug delivery to the lungs, gastrointestinal (GI) tract, and room using a Circulaire nebulizer system. (Drawn from data in Reference 83.)




Fig. 8. Schematic representation of the function of a breath-enhanced nebulizer. Courtesy of Pari Respiratory Equipment.

the nebulizer,^{93,101,104,106} use of a nebulizer and infusion pump (Fig. 10),^{94,96,99,102,103,108} and use of a large-volume nebulizer.^{95,97,98,106,109} Berlinski et al¹¹⁰ reported a consistent and adequate aerosol production by a large-volume nebulizer over a 4-hour period of operation. Reisner et al,¹¹¹ however, reported a more consistent aerosol delivery with a small-volume nebulizer attached to an infusion pump than with a large-volume nebulizer. A commonly used large-volume nebulizer for this therapy is the High-output Extended Aerosol Respiratory Therapy (HEART) nebulizer. Raabe et al¹¹² reported a detailed evaluation of the performance of the HEART nebulizer. At a flow of 10–15 L/min, the aerosol output was 38–50 μ L of aerosolized drug per liter of gas flow, and the solution output was 30–56 mL/hr.

McPeck et al¹¹³ compared the HEART nebulizer to a conventional small-volume nebulizer in a model of adult and pediatric breathing. With the adult breathing pattern they reported similar aerosol delivery from the HEART nebulizer and small-volume nebulizer. For the pediatric breathing pattern the aerosol delivery from the small-volume nebulizer was greater than from the HEART. Both Raabe et al¹¹² and McPeck et al¹¹³ reported an MMAD of about 2 μ m with the HEART nebulizer. An important finding of McPeck et al¹¹³ was that the albuterol delivery from the HEART nebulizer was significantly less than the target dose from the manufacturer's recommended setup.

Nebulizers for Specific Applications

Specially constructed small-volume nebulizers should be used when contamination of the ambient environment with the aerosolized drug needs to be avoided.^{35,114,115} The most common example is aerosolized pentamidine.¹¹⁶ The nebulizer is fitted with one-way valves and filters to prevent gross contamination of the environment. Examples of these devices include the Cadema Aero-Tech II and the Respigard II. These devices produce a very small particle size, with an MMAD of about 1–2 μ m, which is necessary to improve alveolar deposition of the drug.

The Small-Particle Aerosol Generator is used specifically to aerosolize ribavirin (Virazole).^{117–119} The device consists of a nebulizer and a drying chamber. The drying chamber reduces the MMAD of particles to about 1.3 μ m. There are concerns about the potential adverse effects of this drug on health care workers when ribavirin is used. For this reason, a scavenging system should be used when ribavirin is administered.^{120–122} This is a double-enclosure system, with a ribavirin administration hood or mask inside a tent. Two high-flow vacuum scavenging systems




Fig. 9. Tobramycin output from a conventional nebulizer (Hudson) and a breath-enhanced nebulizer (Pari) with changes in inspiratory flow. RF = respirable fraction. (From Reference 31, with permission.)




Fig. 10. System for continuous bronchodilator administration using a conventional nebulizer and an infusion pump. (From Reference 103, with permission.)

aspirate ribavirin from the system through high-efficiency particulate air filters.

Ultrasonic Nebulizers

Ultrasonic nebulizers have been clinically available since the 1960s.^{123,124} Small-volume ultrasonic nebulizers are

commercially available for delivery of inhaled bronchodilators.¹²⁵⁻¹³¹ Although several studies reported greater bronchodilator response with ultrasonic nebulizers than with other aerosol generators,^{127,128} this has not been confirmed in other studies.¹²⁹⁻¹³¹ Large-volume ultrasonic nebulizers are used to deliver inhaled antibiotics in patients with cys-

NEBULIZERS: PRINCIPLES AND PERFORMANCE

Table 3. Summary of Studies Reporting Use of Continuous Nebulization

Author	Study Population	Drug	Research Design	Continuous Nebulizer Design	Major Finding
Portnoy ⁹³	Children with severe asthma	Terbutaline; 1–12 mg/h	Case series	Frequent refilling of conventional nebulizer	Continuously nebulized terbutaline safe and effective for treatment of severe asthma.
Moler ¹⁰³	Children with severe asthma	Terbutaline; 4 mg/h	Case series	Conventional nebulizer with infusion pump	Continuously nebulized terbutaline is an effective therapy for severe asthma.
Calcacone ⁹⁵	Adults with acute asthma	Albuterol; 10 mg over 2 h	Randomized controlled trial	Large volume nebulizer	Continuous nebulization was as effective as intermittent nebulization.
Chippis ¹⁰⁹	Children with bronchospasm	Terbutaline; 4 mg/h	Case series	HEART nebulizer	18 of 23 cases showed significant improvement with continuous nebulization.
Papo ⁹⁶	Children with status asthma	Albuterol; 0.3 mg/kg/h	Randomized controlled trial	Conventional nebulizer with infusion pump	Continuous nebulization was safe and resulted in more rapid improvement than intermittent nebulization.
Lin ⁹⁷	Adults with acute asthma	Albuterol; 30 mg over 110 min	Randomized controlled trial	HEART nebulizer	Continuous nebulization most beneficial in patients with FEV ₁ < 50% predicted.
Rudnitsky ⁹⁸	Adults with acute asthma	Albuterol; 10 mg over 70 min	Randomized controlled trial	HEART nebulizer	Continuous nebulization may be of benefit for patients with peak flow < 200 L/min.
Lin ⁹⁹	Adults with acute asthma	Albuterol; 0.4 mg/kg/h for 4 h	Case series	Conventional nebulizer with infusion pump	High dose continuous nebulization can result in markedly elevated serum albuterol levels and potential cardiac stimulation.
Katz ¹⁰⁰	Infants and children with bronchospasm	Albuterol; 3 mg/kg/h	Case series	Not reported	Continuous albuterol safe and without significant evidence of cardiotoxicity.
Olshaker ¹⁰¹	Adults with acute asthma	Albuterol; 7.5 mg over 1 h	Case series	Frequent refilling of conventional nebulizer	Continuous nebulization was safe and effective.
Baker ¹⁰⁵	Adults with acute asthma	Albuterol; 10 mg/h	Retrospective case control	Conventional nebulizer with infusion pump	Continuous and intermittent nebulization were similar in terms of safety, morbidity, and mortality.
Reisner ¹¹¹	Adults with acute asthma	Albuterol; 7.5 mg/h	Randomized controlled trial	Conventional nebulizer with infusion pump	Continuous nebulization was as safe and effective as intermittent nebulization.
Moler ¹⁰³	Children with acute asthma	Terbutaline; 16 mg over 8 h	Randomized controlled trial	Conventional nebulizer with infusion pump	Continuous nebulization produced similar plasma terbutaline levels and cardiovascular effects as intermittent nebulization.
Shrestha ¹⁰⁴	Adults with acute asthma	Albuterol; 2.5 mg or 7.5 mg over 2 h	Randomized controlled trial	Frequent refilling of conventional nebulizer	The standard dose continuous nebulization group had the greatest improvement with the fewest side effects.
Weber ¹⁰⁶	Adults with acute asthma	Albuterol at 10 mg/h; ipratropium at 1 mg/h	Randomized controlled trial	HEART nebulizer	There was no significant difference in outcomes for patients receiving continuous albuterol alone or continuous albuterol with ipratropium.

Table 4. Advantages and Disadvantages of Ultrasonic Nebulizers

<i>Advantages</i>
Little patient coordination required
Small dead volume
Quiet
Aerosol accumulates during exhalation
High doses possible
No chlorofluorocarbon release
Fast drug delivery
<i>Disadvantages</i>
Expensive
Contamination possible
Prone to electrical and mechanical breakdown
Not all drug formulations available
Drug preparation required

tic fibrosis (eg, tobramycin).¹³²⁻¹³⁵ Ultrasonic nebulizers have also been used during mechanical ventilation,¹³⁶⁻¹³⁹ where they have an advantage in that they do not augment tidal volume, as occurs with pneumatic nebulizers.

Table 4 lists advantages and disadvantages of ultrasonic nebulizers for medication delivery. Table 5 lists factors affecting output from ultrasonic nebulizers.¹⁴⁰ A potential issue with the use of ultrasonic nebulizers is the possibility of drug inactivation by the ultrasonic waves,¹⁴¹ although this has not been shown to occur with commonly-used nebulized medications.

The ultrasonic nebulizer uses a piezoelectric transducer to produce ultrasonic waves that pass through the solution and aerosolize it at the surface of the solution. The ultrasonic nebulizer creates particle sizes of about 1–6 μm MMAD, depending on the manufacturer of the device.¹⁴⁰ The volume output of the ultrasonic nebulizer is about 1–6 mL/min, depending on the manufacturer of the device.¹⁴⁰

An ultrasonic nebulizer has 3 components: the power unit, the transducer, and a fan. The power unit converts electrical energy to high-frequency ultrasonic waves at a frequency of 1.3–2.3 megahertz.¹⁴⁰ The frequency of the ultrasonic waves determines the size of the particles, with an inverse relationship between frequency and particle size. The frequency is not user adjustable. The power unit also controls the amplitude of the ultrasonic waves. This is user adjustable, with an increase in amplitude resulting in an increase in output from the ultrasonic nebulizer. The trans-

ducer vibrates at the frequency of the ultrasonic waves applied to it (piezoelectric effect). The transducer is found in two shapes, concave (focused) and flat (unfocused).¹⁴⁰ Concave transducers produce a higher output but require a constant level of solution for proper operation. The conversion of ultrasonic energy to mechanical energy by the transducer produces heat, which is absorbed by the solution over the transducer.

In some ultrasonic nebulizers, the solution to be nebulized is placed directly over the transducer. In others, the solution to be nebulized is placed into a nebulization chamber and a water couplant chamber is placed between the transducer and the medication chamber. A fan is used to deliver the aerosol produced by the ultrasonic nebulizer to the patient, or the aerosol is evacuated from the nebulization chamber by the inspiratory flow of the patient.

Summary

Nebulizers have been used clinically for many years. Despite the increasing use of metered-dose inhalers and dry powder inhalers, it is likely that nebulizers will continue to be used in selected patients. A number of factors affect nebulizer performance, and these should be appreciated by clinicians who use these devices. Several new designs have recently become available that improve the performance of the nebulizer, but their cost-effectiveness remains to be determined.

REFERENCES

1. AARC Clinical Practice Guideline. Selection of aerosol delivery device. American Association for Respiratory Care. *Respir Care* 1992;37(8):891–897.
2. O'Donohue WJ Jr. Guidelines for the use of nebulizers in the home and at domiciliary sites: report of a consensus conference. NAM-DRC Consensus Group. *Chest* 1996;109(3):814–820.
3. Turner MO, Gafni A, Swan D, FitzGerald JM. A review and economic evaluation of bronchodilator delivery methods in hospitalized patients. *Arch Intern Med* 1996;156(18):2113–2118.
4. Cates CJ. Holding chambers versus nebulizers for beta-agonist treatment of acute asthma. The Cochrane Database or Systematic Reviews, 1999.
5. Dalby RN, Tiano SL, Hickey AJ. Medical devices for the delivery of therapeutic aerosols to the lungs. In: Hickey AJ, editor. *Inhalation aerosols: physical and biological basis for therapy*. New York: Marcel Dekker; 1996: 441–473.
6. Newman SP. Aerosol generators and delivery systems. *Respir Care* 1991;36(9):939–951.
7. Dennis JH, Hendrick DJ. Design characteristics for drug nebulizers. *J Med Eng Technol* 1992;16(2):63–68.
8. Smye SW, Jollie MI, Littlewood JM. A mathematical model of some aspects of jet nebuliser performance. *Clin Phys Physiol Meas* 1991;12(3):289–300.
9. Niven RW, Brain JD. Some functional characteristics of air-jet nebulizers. *Int J Pharm* 1994;104:73–85.
10. Nerbrink O, Dahlback M, Hansson HC. Why do medical nebulizers differ in their output and particle size characteristics? *J Aerosol Med* 1994;7:259–276.

Table 5. Factors Affecting Output from Ultrasonic Nebulizers

Fluid characteristics: density, viscosity, surface tension, vapor pressure
Piezoelectric transducer: frequency of vibration, amplitude of vibration, configuration (focused or flat)
Coupling of medication chamber to transducer
Medication chamber: size, baffles
Flow from fan

11. Fuller HD, Dolovich MB, Chambers C, Newhouse MT. Aerosol delivery during mechanical ventilation: a predictive in vitro lung model. *J Aerosol Med* 1992;5:251-259.
12. AARC Clinical Practice Guideline. Selection of device, administration of bronchodilator, and evaluation of response to therapy in mechanically ventilated patients. American Association for Respiratory Care. *Respir Care* 1999;44(1):105-113.
13. Diot P, Morra L, Smaldone GC. Albuterol delivery in a model of mechanical ventilation: comparison of metered-dose inhaler and nebulizer efficiency. *Am J Respir Crit Care Med* 1995;152(4 Pt 1):1391-1394.
14. O'Riordan TG, Greco MJ, Perry RJ, Smaldone GC. Nebulizer function during mechanical ventilation. *Am Rev Respir Dis* 1992;145(5):1117-1122.
15. Niven RW. Atomization and nebulizers. In: Hickey AJ, editor. *Inhalation aerosols: physical and biological basis for therapy*. New York: Marcel Dekker; 1996: 273-312.
16. Everard ML, Evans M, Milner AD. Is tapping jet nebulisers worthwhile? *Arch Dis Child* 1994;70(6):538-539.
17. Malone RA, Hollie MC, Glynn-Barnhart A, Nelson HS. Optimal duration of nebulized albuterol therapy. *Chest* 1993;104(4):1114-1118.
18. Phipps PR, Gonda I. Droplets produced by medication nebulizers: some factors affecting their size and solute concentration. *Chest* 1990;97(6):1327-1332.
19. Wood JA, Wilson RSE, Bray C. Changes in salbutamol concentration in the reservoir solution of a jet nebulizer. *Br J Dis Chest* 1986;80(2):164-169.
20. Stapleton KW, Finlay WH. Determining solution concentration within aerosol droplets output by jet nebulizers. *J Aerosol Sci* 1995;26:137-145.
21. O'Callaghan C, Clarke AR, Milner AD. Inaccurate calculation of drug output from nebulisers. *Eur J Pediatr* 1989;148(5):473-474.
22. Tandon R, McPeck M, Smaldone GC. Measuring nebulizer output: aerosol production vs gravimetric analysis. *Chest* 1997;111(5):1361-1365.
23. Clay MM, Pavia D, Newman SP, Lennard-Jones T, Clarke SW. Assessment of jet nebulisers for lung aerosol therapy. *Lancet* 1983;2(8350):592-594.
24. Aerosol consensus statement—1991. American Association for Respiratory Care. *Respir Care* 1991;36(9):916-921.
25. Hess D, Fisher D, Williams P, Pooler S, Kacmarek RM. Medication nebulizer performance: effects of diluent volume, nebulizer flow, and nebulizer brand. *Chest* 1996;110(2):498-505.
26. Loffert DT, Ikle D, Nelson HS. A comparison of commercial jet nebulizers. *Chest* 1994;106(6):1788-1792.
27. Alvine GF, Rodgers P, Fitzsimmons KM, Ahrens RC. Disposable jet nebulizers. How reliable are they? *Chest* 1992;101(2):316-319.
28. Hollie MC, Malone RA, Skufca RM, Nelson HS. Extreme variability in aerosol output of the DeVilbiss 646 jet nebulizer. *Chest* 1991;100(5):1339-1344.
29. Langford SA, Allen MB. Salbutamol output from two jet nebulizers. *Respir Med* 1993;87(2):99-103.
30. Ho SL, Coates AL. Effect of dead volume on the efficiency and the cost to deliver medications in cystic fibrosis with four disposable nebulizers. *Can Respir J* 1999;6(3):253-260.
31. Coates AL, MacNeish CF, Lands LC, Meisner D, Kelemen S, Vadas EB. A comparison of the availability of tobramycin for inhalation from vented vs unvented nebulizers. *Chest* 1998;113(4):951-956.
32. Arossa W, Quagliotti F, Sala M, Spinaci S, De Candussio G. Different performance of two commercial nebulizers. *Respiration* 1984;46(1):128-132.
33. Hess D, Horney D, Snyder T. Medication-delivery performance of eight small-volume, hand-held nebulizers: effects of diluent volume, gas flowrate, and nebulizer model. *Respir Care* 1989;34(8):717-723.
34. Hurley PK, Smye SW, Cunliffe H. Assessment of antibiotic aerosol generation using commercial jet nebulizers. *J Aerosol Med* 1994;7(3):217-228.
35. Smaldone GC, Perry RJ, Deutsch DG. Characteristics of nebulizers used in the treatment of AIDS-related pneumocystis carinii pneumonia. *J Aerosol Med* 1988;1:113-126.
36. Waldrep JC, Keyhani K, Black M, Knight V. Operating characteristics of 18 different continuous-flow jet nebulizers with beclomethasone dipropionate liposome aerosol. *Chest* 1994;105(1):106-110.
37. Hartley-Sharp CJ, Booth H, Johns DP, Walters EH. Differences in aerosol output and airways responsiveness between the DeVilbiss 40 and 45 hand held nebulisers. *Thorax* 1995;50(6):635-638.
38. MacNeish CF, Meisner D, Thibert R, Kelemen S, Vadas EB, Coates AL. A comparison of pulmonary availability between Ventolin (albuterol) nebulizer and Ventolin (albuterol) Respirator Solution. *Chest* 1997;111(1):204-208.
39. Coates AL, MacNeish CF, Meisner D, Kelemen S, Thibert R, MacDonald J, Vadas E. The choice of jet nebulizer, nebulizing flow, and addition of albuterol affects the output of tobramycin aerosols. *Chest* 1997;111(5):1206-1212.
40. Newman SP, Woodman G, Clarke SW. Deposition of carbenicillin aerosols in cystic fibrosis: effects of nebulizer system and breathing pattern. *Thorax* 1988;43(4):318-322.
41. O'Doherty M, Thomas S, Page C, Bradbeer C, Nunan T, Bateman N. Pulmonary deposition of nebulised pentamidine isethionate: effect of nebuliser type, dose, and volume of fill. *Thorax* 1990;45(6):460-464.
42. Hardy JG, Newman SP, Knoch M. Lung deposition from four nebulizers. *Respir Med* 1993;87(6):461-465.
43. Johnson MA, Newman SP, Bloom R, Talaei N, Clarke SW. Delivery of albuterol and ipratropium bromide from two nebulizer systems in chronic stable asthma: efficacy and pulmonary deposition. *Chest* 1989;96(1):6-10.
44. Standaert TA, Morlin GL, Williams-Warren J, Joy P, Pepe MS, Weber A, Ramsey BW. Effects of repetitive use and cleaning techniques of disposable jet nebulizers on aerosol generation. *Chest* 1998;114(2):577-586.
45. Kradjan WA, Lakshminarayan S. Efficiency of air compressor-driven nebulizers. *Chest* 1985;87(4):512-516.
46. Hadfield JW, Windebank WJ, Bateman JRM. Is driving gas flow rate clinically important for nebulizer therapy? *Br J Dis Chest* 1986;80(1):50-54.
47. Clay MM, Pavia D, Newman SP, Clarke SW. Factors influencing the size distribution of aerosols from jet nebulisers. *Thorax* 1983;38(10):755-759.
48. Everard ML, Clark AR, Milner AD. Drug delivery from jet nebulisers. *Arch Dis Child* 1992;67(5):586-591.
49. Flament MP, Leterme P, Burnouf T, Gayot A. Influence of formulation on jet nebulisation quality of alpha 1 protease inhibitor. *Int J Pharm* 1999;178(1):101-109.
50. Hess DR, Acosta FL, Ritz RH, Kacmarek RM, Camargo CA Jr. The effect of heliox on nebulizer function using a beta-agonist bronchodilator. *Chest* 1999;115(1):184-189.
51. Martonen T, Yang Y. Deposition mechanics of pharmaceutical particles in human airways. In: Hickey AJ, editor. *Inhalation aerosols: physical and biological basis for therapy*. New York: Marcel Dekker; 1996: 3-27.
52. Smaldone GC. Drug delivery via aerosol systems: concept of "aerosol inhaled". *J Aerosol Med* 1991;4(3):229-235.

53. Steventon RD, Wilson RSE. Facemask or mouthpiece for delivery of nebulized bronchodilator aerosols? *Br J Dis Chest* 1981;75(1):88-90.
54. Everard ML, Hardy JG, Milner AD. Comparison of nebulised aerosol deposition in the lungs of healthy adults following oral and nasal inhalation. *Thorax* 1993;48(10):1045-1046.
55. Lowenthal D, Kattan M. Facemasks versus mouthpieces for aerosol treatment of asthmatic children. *Pediatr Pulmonol* 1992;14(3):192-196.
56. Smedsaas-Lofvenberg A, Nilsson K, Moa G, Axelsson I. Nebulization of drugs in a nasal CPAP system. *Acta Paediatr* 1999;88(1):89-92.
57. Parkes SN, Bersten AD. Aerosol kinetics and bronchodilator efficacy during continuous positive airway pressure delivered by face mask. *Thorax* 1997;52(2):171-175.
58. Kairaitis K, Garlick SR, Wheatley JR, Amis TC. Route of breathing in patients with asthma. *Chest* 1999;116(6):1646-1652.
59. Melchor R, Biddiscombe MF, Mak VHF, Short MD, Spiro SG. Lung deposition patterns of directly labelled salbutamol in normal subjects and in patients with reversible airflow obstruction. *Thorax* 1993;48(5):506-511.
60. Pavia D, Thompson ML, Clarke SW, Shannon HS. Effect of lung function and mode of inhalation on penetration of aerosol into the human lung. *Thorax* 1977;32(2):194-197.
61. Lipworth BJ, Clarke DJ. Effects of airway calibre on lung delivery of nebulised salbutamol. *Thorax* 1997;52(12):1036-1039.
62. Swift DL, Carpin JC, Mitzner W. Pulmonary penetration and deposition of aerosols in different gases: fluid flow effects. *Ann Occup Hyg* 1982;26(1-4):109-117.
63. Svartengren M, Anderson M, Philipson K, Camner P. Human lung deposition of particles suspended in air or in helium/oxygen mixture. *Exp Lung Res* 1989;15(4):575-585.
64. Esch JL, Specktor DM, Lippmann M. Effect of lung airway branching pattern and gas composition on particle deposition: II. Experimental studies in human and canine lungs. *Exp Lung Res* 1988;14(3):321-348.
65. Anderson M, Svartengren M, Bylin G, Philipson K, Camner P. Deposition in asthmatics of particles inhaled in air or in helium-oxygen. *Am Rev Respir Dis* 1993;147(3):524-528.
66. Anderson M, Svartengren M, Philipson K, Camner P. Deposition in man of particles suspended in air or in helium-oxygen mixture at different flow rates. *J Aerosol Med* 1990;3:209-216.
67. Henderson SO, Acharya P, Kilaghbian T, Perez J, Korn CS, Chan LS. Use of heliox-driven nebulizer therapy in the treatment of acute asthma. *Ann Emerg Med* 1999;33(2):141-146.
68. Pollack CV Jr, Fleisch KB, Dowsey K. Treatment of acute bronchospasm with beta-adrenergic agonist aerosols delivered by a nasal bilevel positive airway pressure circuit. *Ann Emerg Med* 1995;2(65):552-557.
69. Dhand R, Tobin MJ. Bronchodilator delivery with metered-dose inhalers in mechanically ventilated patients. *Eur Respir J* 1996;9:585-595.
70. Dhand R, Tobin MJ. Inhaled bronchodilator therapy in mechanically-ventilated patients (review). *Am J Respir Crit Care Med* 1997;156(1):3-10.
71. Mantous CA, Hall JB. Administration of therapeutic aerosols to mechanically ventilated patients (review). *Chest* 1994;106(2):560-571.
72. Hess D. Inhaled bronchodilators during mechanical ventilation: delivery techniques, evaluation of response, and cost-effectiveness. *Respir Care* 1994;39(2):105-122.
73. Fink JB, Tobin MJ, Dhand R. Bronchodilator therapy in mechanically ventilated patients. *Respir Care* 1999;44(1):53-69.
74. Craven DE, Lichtenberg DA, Goularte TA, Make BJ, McCabe WR. Contaminated medication nebulizers in mechanical ventilator circuits: source of bacterial aerosols. *Am J Med* 1984;77(5):834-838.
75. Beaty CD, Ritz RH, Benson MS. Continuous in-line nebulizers complicate pressure support ventilation. *Chest* 1989;96(6):1360-1363.
76. Hanhan U, Kissoon N, Payne M, Taylor C, Murphy S, DeNicola LK. Effects of in-line nebulization on preset ventilatory variables. *Respir Care* 1993;38(5):474-478.
77. Marik P, Hogan J, Krikorian J. A comparison of bronchodilator therapy delivered by nebulization and metered-dose inhaler in mechanically ventilated patients. *Chest* 1999;115(6):1653-1657.
78. Dennis JH. A review of issues relating to nebulizer standards. *J Aerosol Med* 1998;11 Suppl 1:S73-S77.
79. Pisut FM. Comparison of medication delivery by T-nebulizer with inspiratory and expiratory reservoir. *Respir Care* 1989;34(11):985-988.
80. Marshall LM, Francis PW, Khafagi FA. Aerosol deposition in cystic fibrosis using an aerosol conservation device and a conventional jet nebulizer. *J Paediatr Child Health* 1994;30(1):65-67.
81. Thomas SH, Langford JA, George RDG, Geddes DM. Improving the efficacy of drug administration with jet nebulizers (letter). *Lancet* 1988;1(8577):126.
82. Piper SD. In vitro comparison of the Circulaire and AeroTee to a traditional nebulizer T-Piece with corrugated tubing. *Respir Care* 2000;45(3):313-319.
83. Mason JW, Miller WC, Small S. Comparison of aerosol delivery via Circulaire system vs conventional small volume nebulizer. *Respir Care* 1994;39(12):1157-1161.
84. Mason JW, Miller WC. Comparison of aerosol delivery via circulaire nebulizer system versus a disposable nebulizer in COPD patients. *Respir Care* 1996;41(11):1006-1008.
85. Hoffman L, Smithline H. Comparison of Circulaire to conventional small volume nebulizer for the treatment of bronchospasm in the emergency department. *Respir Care* 1997;42(12):1170-1174.
86. Newnham DM, Lipworth BJ. Nebuliser performance, pharmacokinetics, airways and systemic effects of salbutamol given via a novel nebuliser system ("Venstream"). *Thorax* 1994;49(8):762-770.
87. Newman SP, Pitcairn GR, Hooper G, Knoch M. Efficient drug delivery to the lungs from a continuously operated open-vent nebulizer and low pressure compressor system. *Eur Respir J* 1994;7(6):1177-1181.
88. Devadason SG, Everard ML, Linto JM, Le Souef PN. Comparison of drug delivery from conventional versus "Venturi" nebulizers. *Eur Respir J* 1997;10(11):2479-2483.
89. Barry PW, O'Callaghan C. The output of budesonide from nebulizers. *J Allergy Clin Immunol* 1998;102(2):321-322.
90. Suez DS, Chai H. A standard method of intermittent inhaled therapy via a jet nebulizer. *Ann Allergy* 1986;57(4):245-248.
91. McPeck M, O'Riordan TG, Smaldone GC. Choice of mechanical ventilator: influence on nebulizer performance. *Respir Care* 1993;38(8):887-895.
92. Hughes JM, Saez J. Effects of nebulizer mode and position in a mechanical ventilator circuit on dose efficiency. *Respir Care* 1987;32(12):1131-1135.
93. Portnoy J, Aggarwal J. Continuous terbutaline nebulization for the treatment of severe exacerbations of asthma in children. *Ann Allergy* 1988;60(4):368-371.
94. Moler FW, Hurwitz ME, Custer JR. Improvement in clinical asthma score and PaCO₂ in children with severe asthma treated with continuously nebulized terbutaline. *J Allergy Clin Immunol* 1988;81(6):1101-1109.

95. Calcacone A, Wolkove N, Stern E, Afilalo M, Rosental TM, Kreisman H. Continuous nebulization of albuterol (salbutamol) in acute asthma. *Chest* 1990;97(3):693-697.
96. Papo MC, Frank J, Thompson AE. A prospective, randomized study of continuous versus intermittent nebulized albuterol for severe status asthmaticus in children. *Crit Care Med* 1993;21(10):1479-1486.
97. Lin RY, Sauter D, Newman T, Sirleaf J, Walters J, Tavakol M. Continuous versus intermittent albuterol nebulization in the treatment of acute asthma. *Ann Emerg Med* 1993;22(12):1847-1853.
98. Rudnitsky GS, Eberlein RS, Schoffstall JM, Mazur JE, Spivey WH. Comparison of intermittent and continuously nebulized albuterol for treatment of asthma in an urban emergency department. *Ann Emerg Med* 1993;22(12):1842-1846.
99. Lin RY, Smith AJ, Hergenroeder P. High serum albuterol levels and tachycardia in adult asthmatics treated with high-dose continuously aerosolized albuterol. *Chest* 1993;103(1):221-225.
100. Katz RW, Kelly HW, Crowley MR, Grad R, McWilliams BC, Murphy SJ. Safety of continuously nebulized albuterol for bronchospasm in infants and children. *Pediatrics* 1993;92(5):666-669.
101. Olshaker J, Jerrard D, Barish RA, Brandt G, Hooper F. The efficacy and safety of a continuous albuterol protocol for the treatment of acute adult asthma attacks. *Am J Emerg Med* 1993;11(2):131-133.
102. Reisner C, Kotch A, Dworkin G. Continuous versus frequent intermittent nebulization of albuterol in acute asthma: a randomized, prospective study. *Ann Allergy Asthma Immunol* 1995;75(1):41-47.
103. Moler FW, Johnson CE, Van Laanen C, Palmisano JM, Nasr SZ, Akingbola O. Continuous versus intermittent nebulized terbutaline: plasma levels and effects. *Am J Respir Crit Care Med* 1995;151(3 Pt 1):602-606.
104. Shrestha M, Bidadi K, Gourlay S, Hayes J. Continuous vs intermittent albuterol, at high and low doses, in the treatment of severe acute asthma in adults. *Chest* 1996;110(1):42-47.
105. Baker EK, Willsie SK, Marinac JS, Salzman GA. Continuously nebulized albuterol in severe exacerbations of asthma in adults: a case controlled study. *J Asthma* 1997;34:521-530.
106. Weber EJ, Levitt MA, Covington JK, Gambrioli E. Effect of continuously nebulized ipratropium bromide plus albuterol on emergency department length of stay and hospital admission rates in patients with acute bronchospasm: a randomized, controlled trial. *Chest* 1999;115(4):937-944.
107. Portnoy J, Nadel G, Amado M, Willsie-Ediger S. Continuous nebulization for status asthmaticus (review). *Ann Allergy* 1992;69(1):71-79.
108. Voss KR, Willsie-Ediger SK, Pyszczyńska DR, Nelson KA. Description of a delivery method for continuously aerosolized albuterol in status asthmaticus. *J Asthma* 1990;27(1):37-39.
109. Chipp BE, Blackney DA, Black LE, Moody RR, Marino JT, Ridell RC, Wong GA. Vortran high output extended aerosol respiratory therapy (HEART) for delivery of continuously nebulized terbutaline for the treatment of acute bronchospasm. *Pediatr Asthma Allergy Immunol* 1990;471-277.
110. Berlinski A, Waldrep JC. Four hours of continuous albuterol nebulization. *Chest* 1998;114(3):847-853.
111. Reisner C, Lee J, Kotch A, Dworkin G. Comparison of volume output from two different continuous nebulizer systems. *Ann Allergy Asthma Immunol* 1996;76(2):209-213.
112. Raabe OG, Wong TM, Wong GB, Roxburgh JW, Piper SD, Lee JJ. Continuous nebulization therapy for asthma with aerosols of beta2 agonists. *Ann Allergy Asthma Immunol* 1998;80(6):499-508.
113. McPeck M, Tandon R, Hughes K, Smaldone GC. Aerosol delivery during continuous nebulization. *Chest* 1997;111(5):1200-1205.
114. Corkery KJ, Luce JM, Montgomery AB. Aerosolized pentamidine for treatment and prophylaxis of *Pneumocystis carinii* pneumonia: an update. *Respir Care* 1988;33(8):676-685.
115. Vinciguerra C, Smaldone G. Treatment time and patient tolerance for pentamidine delivery by Respigard II and AeroTech II. *Respir Care* 1990;35(11):1037-1041.
116. Matthys H, Herczeg R. Dosing strategies for aerosol delivery to the lung parenchyma, with specific recommendations for pentamidine. *Respir Care* 1991;36(9):989-993.
117. Demers RR, Parker J, Frankel LR, Smith DW. Administration of ribavirin to neonatal and pediatric patients during mechanical ventilation. *Respir Care* 1986;31(12):1188-1195.
118. American Academy of Pediatrics Committee on Infectious Diseases. Use of ribavirin in the treatment of respiratory syncytial virus infection. *Pediatrics* 1993;92(3):501-504.
119. Smith DW, Frankel LR, Mathers LH, Tang AT, Ariagno RL, Prober CG. A controlled trial of aerosolized ribavirin in infants receiving mechanical ventilation for severe respiratory syncytial virus infection. *N Engl J Med* 1991;325(1):24-29.
120. Kacmarek RM, Kratochvil J. Evaluation of a double-enclosure double-vacuum unit scavenging system for ribavirin administration. *Respir Care* 1992;37(1):37-45.
121. Charney W, Corkery KJ, Kraemer R, Wugofski L. Engineering and administrative controls to contain aerosolized ribavirin: results of simulation and application to one patient. *Respir Care* 1990;35(11):1042-1048.
122. Wahlin B, Malmstrom B, Soop M, Hellstrom LG. A pediatric canopy system for aerosol administration and minimized environmental pollution. *Acta Anaesthesiol Scand* 1996;40(8 Pt 1):932-939.
123. Stevens HR, Albregt HB. Assessment of ultrasonic nebulization. *Anesthesiology* 1966;27(5):648-653.
124. Modell JH, Giammona ST, Davis JH. Effect of chronic exposure to ultrasonic aerosols on the lung. *Anesthesiology* 1967;28(4):680-688.
125. Dennis JH, Stenton SC, Beach JR, Avery AJ, Walters EH, Hendrick DJ. Jet and ultrasonic nebuliser output: use of a new method for direct measurement of aerosol output. *Thorax* 1990;45(10):728-732.
126. Newman SP, Pellow GD, Clarke SW. In vitro comparison of DeVilbiss jet and ultrasonic nebulizers. *Chest* 1987;92(6):991-994.
127. Fok TF, Lam K, Ng PC, So HK, Cheung KL, Wong W, So KW. Randomised crossover trial of salbutamol aerosol delivered by metered-dose inhaler, jet nebuliser, and ultrasonic nebuliser in chronic lung disease. *Arch Dis Child Fetal Neonatal Ed* 1998;79(2):F100-F104.
128. Ballard RD, Bogin RM, Pak J. Assessment of bronchodilator response to a beta-adrenergic delivered from an ultrasonic nebulizer. *Chest* 1991;100(2):410-415.
129. Nakanishi AK, Lamb BM, Foster C, Rubin BK. Ultrasonic nebulization of albuterol is no more effective than jet nebulization for the treatment of acute asthma in children. *Chest* 1997;111(6):1505-1508.
130. Larsen KR, Svendsen UG, Molgaard F, Petersen BN. Comparability of albuterol delivered by a piezoelectric device versus metered-dose inhaler in patients with chronic obstructive airways disease. *J Aerosol Med* 1998;11(2):81-88.
131. Olivenstein R, Wolkove N, Cohen C, Frank H, Kreisman H. A comparison of responses to albuterol delivered by two aerosol devices. *Chest* 1986;90(3):392-395.
132. Eisenberg J, Pepe M, Williams-Warren J, Vasiliev M, Montgomery AB, Smith AL, Ramsey BW. A comparison of peak sputum tobramycin concentration in patients with cystic fibrosis using jet and ultrasonic nebulizer systems. Aerosolized Tobramycin Study Group. *Chest* 1997;111(4):955-962.

133. O'Riordan TG, Amram JC. Effect of nebulizer configuration on delivery of aerosolized tobramycin. *J Aerosol Med* 1997;10(1):13-23.
134. Weber A, Morlin G, Cohen M, Williams-Warren J, Ramsey B, Smith A. Effect of nebulizer type and antibiotic concentration on device performance. *Pediatr Pulmonol* 1997;23(4):249-260.
135. Weber A, Smith A, Williams-Warren J, Ramsey B, Covert DS. Nebulizer delivery of tobramycin to the lower respiratory tract. *Pediatr Pulmonol* 1994;17(5):331-339.
136. Harvey CJ, O'Doherty MJ, Page CJ, Thomas SHL, Nunan TO, Treacher DF. Comparison of jet and ultrasonic nebulizer pulmonary aerosol deposition during mechanical ventilation. *Eur Respir J* 1997;10(4):905-909.
137. Thomas SH, O'Doherty MJ, Page CJ, Treacher DF, Nunan TO. Delivery of ultrasonic nebulized aerosols to a lung model during mechanical ventilation. *Am Rev Respir Dis* 1993;148(4 Pt 1):872-877.
138. Kemming GI, Kreyling W, Habler O, Merkel M, Kleen M, Welte M, et al. Aerosol production and aerosol droplet size distribution during mechanical ventilation (IPPV) with a new ultrasonic nebulizer. *Eur J Med Res* 1996;1(7):321-327.
139. Williams L, Fletcher GC, Daniel M, Kinsella J. A simple in vitro method for the evaluation of an ultrasonic nebulizer for drug delivery to intubated, ventilated patients and the effect of nebulizer and ventilator settings on the uptake of fluid from the nebulizer chamber. *Eur J Anaesthesiol* 1999;16(7):479-484.
140. Greenspan BJ. Ultrasonic and electrohydrodynamic methods for aerosol generation. In: Hickey AJ, editor. *Inhalation aerosols: physical and biological basis for therapy*. New York: Marcel Dekker; 1996: 313-335.
141. Gale AE. Drug degeneration during ultrasonic nebulization. *J Aerosol Sci* 1985;16:265.